

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: BAPS03
 Lab Code: ACE SDG NO.: Q2892
 Lab File ID: VY023162.D Date Analyzed: 08/20/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:26
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	347499	13.347				
UPPER LIMIT	694998	13.847				
LOWER LIMIT	173750	12.847				
EPA SAMPLE NO.						
1-FLOOR-NORTH-EAST-CORNER-BAC	285220	13.35				
1-FLOOR-DRILL-B-BACK	308457	13.34				
BACK-1-FLOOR-CENTER-MIDDLE-C	347720	13.35				
3-FLOOR-NORTH-D	329619	13.34				
3-FLOOR-SOUTH-E	335210	13.35				
FRONT-1-FLOOR-F	312976	13.35				
VY0820SBL01	382880	13.35				
VY0820SBS01	323804	13.35				
VY0820SBSD01	318462	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.