

Continuing Calibration Report for SequenceID : FC082825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069764.D

Aliphatic C9-C12	7254459.000	60.000	3.295	6.591	120907.650	126974.123	4.778
Aliphatic C12-C16	4844525.000	40.000	6.592	9.993	121113.125	128339.940	5.631
Aliphatic C16-C21	6589838.000	60.000	9.994	13.361	109830.633	117758.364	6.732
Aliphatic C21-C28	7581105.000	80.000	13.362	17.027	94763.813	102626.001	7.661
Aliphatic C28-C40	8563720.000	120.000	17.028	21.995	71364.333	81443.743	12.376
Aliphatic EPH	34833647.000	360.000	3.295	21.995	96760.131	105002.433	7.850