

Continuing Calibration Report for SequenceID : FC082825AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069774.D

Aliphatic C9-C12	7110180.000	60.000	3.295	6.591	118503.000	126974.123	6.672
Aliphatic C12-C16	4748910.000	40.000	6.592	9.993	118722.750	128339.940	7.494
Aliphatic C16-C21	6483577.000	60.000	9.994	13.361	108059.617	117758.364	8.236
Aliphatic C21-C28	7500716.000	80.000	13.362	17.027	93758.950	102626.001	8.640
Aliphatic C28-C40	8640237.000	120.000	17.028	21.995	72001.975	81443.743	11.593
Aliphatic EPH	34483620.000	360.000	3.295	21.995	95787.833	105002.433	8.776