

Continuing Calibration Report for SequenceID : FC082925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069788.D

Aliphatic EPH	35745607.000	360.000	3.297	22.011	99293.353	105002.433	5.437
Aliphatic C9-C12	7140730.000	60.000	3.297	6.594	119012.167	126974.123	6.271
Aliphatic C12-C16	4740056.000	40.000	6.595	9.997	118501.400	128339.940	7.666
Aliphatic C16-C21	6608492.000	60.000	9.998	13.367	110141.533	117758.364	6.468
Aliphatic C21-C28	7800374.000	80.000	13.368	17.033	97504.675	102626.001	4.990
Aliphatic C28-C40	9455955.000	120.000	17.034	22.011	78799.625	81443.743	3.247