

Continuing Calibration Report for SequenceID : FD082825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049690.D

Aromatic C10-C12	5034296.000	40.000	4.405	6.149	125857.400	107939.663	-16.600
Aromatic C12-C16	8469044.000	60.000	6.150	8.777	141150.733	123580.152	-14.218
Aromatic C16-C21	11070116.000	80.000	8.778	13.067	138376.450	126486.361	-9.400
Aromatic C21-C36	20876582.000	180.000	13.068	18.488	115981.011	112618.254	-2.986
Aromatic EPH	45450038.000	360.000	4.405	18.488	126250.106	117007.195	-7.899