

Continuing Calibration Report for SequenceID : FD082825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049700.D

Aromatic C10-C12	4555410.000	40.000	4.405	6.149	113885.250	107939.663	-5.508
Aromatic C12-C16	7648803.000	60.000	6.150	8.777	127480.050	123580.152	-3.156
Aromatic C16-C21	10446414.000	80.000	8.778	13.067	130580.175	126486.361	-3.237
Aromatic C21-C36	20477231.000	180.000	13.068	18.488	113762.394	112618.254	-1.016
Aromatic EPH	43127858.000	360.000	4.405	18.488	119799.606	117007.195	-2.387