

Continuing Calibration Report for SequenceID : FD082925AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049703.D

Aromatic C10-C12	5075439.000	40.000	4.405	6.148	126885.975	107939.663	-17.553
Aromatic C12-C16	8650554.000	60.000	6.149	8.777	144175.900	123580.152	-16.666
Aromatic C16-C21	11299431.000	80.000	8.778	13.068	141242.888	126486.361	-11.666
Aromatic C21-C36	21456853.000	180.000	13.069	18.488	119204.739	112618.254	-5.849
Aromatic EPH	46482277.000	360.000	4.405	18.488	129117.436	117007.195	-10.350