

Continuing Calibration Report for SequenceID : FD082925AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049706.D

Aromatic C10-C12	4596448.000	40.000	4.405	6.148	114911.200	107939.663	-6.459
Aromatic C12-C16	7793899.000	60.000	6.149	8.777	129898.317	123580.152	-5.113
Aromatic C16-C21	10420332.000	80.000	8.778	13.068	130254.150	126486.361	-2.979
Aromatic C21-C36	20194414.000	180.000	13.069	18.488	112191.189	112618.254	0.379
Aromatic EPH	43005093.000	360.000	4.405	18.488	119458.592	117007.195	-2.095