

Continuing Calibration Report for SequenceID : FD082925AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049715.D

Aromatic C10-C12	4250758.000	40.000	4.405	6.148	106268.950	107939.663	1.548
Aromatic C12-C16	7092776.000	60.000	6.149	8.777	118212.933	123580.152	4.343
Aromatic C16-C21	9537588.000	80.000	8.778	13.068	119219.850	126486.361	5.745
Aromatic C21-C36	19653849.000	180.000	13.069	18.488	109188.050	112618.254	3.046
Aromatic EPH	40534971.000	360.000	4.405	18.488	112597.142	117007.195	3.769