

Continuing Calibration Report for SequenceID : FE090525AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE055688.D

Aliphatic C9-C12	7917072.000	60.000	3.311	6.945	131951.200	128403.808	-2.763
Aliphatic C12-C16	5625095.000	40.000	6.946	10.397	140627.375	135095.243	-4.095
Aliphatic C16-C21	9027864.000	60.000	10.398	13.775	150464.400	144217.196	-4.332
Aliphatic C21-C28	12129334.000	80.000	13.776	17.446	151616.675	143235.734	-5.851
Aliphatic C28-C40	17741065.000	120.000	17.447	22.456	147842.208	137719.667	-7.350
Aliphatic EPH	52440430.000	360.000	3.311	22.456	145667.861	138184.135	-5.416