

Continuing Calibration Report for SequenceID : FF082925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016272.D

Aliphatic C9-C12	7082171.000	60.000	3.232	6.870	118036.183	118465.299	0.362
Aliphatic C12-C16	4836082.000	40.000	6.871	10.313	120902.050	122693.653	1.460
Aliphatic C16-C21	7617590.000	60.000	10.314	13.684	126959.833	128123.078	0.908
Aliphatic C21-C28	10052655.000	80.000	13.685	17.353	125658.188	125602.733	-0.044
Aliphatic C28-C40	12541251.000	120.000	17.354	22.318	104510.425	101676.216	-2.787
Aliphatic EPH	42129749.000	360.000	3.232	22.318	117027.081	116534.481	-0.423