

Continuing Calibration Report for SequenceID : FF082925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016280.D

Aliphatic C9-C12	6881481.000	60.000	3.232	6.870	114691.350	118465.299	3.186
Aliphatic C12-C16	4702126.000	40.000	6.871	10.313	117553.150	122693.653	4.190
Aliphatic C16-C21	7408221.000	60.000	10.314	13.684	123470.350	128123.078	3.631
Aliphatic C21-C28	9641304.000	80.000	13.685	17.353	120516.300	125602.733	4.050
Aliphatic C28-C40	11365882.000	120.000	17.354	22.318	94715.683	101676.216	6.846
Aliphatic EPH	39999014.000	360.000	3.232	22.318	111108.372	116534.481	4.656