

Continuing Calibration Report for SequenceID : FF082925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016295.D

Aliphatic C9-C12	6818792.000	60.000	3.232	6.870	113646.533	118465.299	4.068
Aliphatic C12-C16	4720737.000	40.000	6.871	10.313	118018.425	122693.653	3.810
Aliphatic C16-C21	7358336.000	60.000	10.314	13.684	122638.933	128123.078	4.280
Aliphatic C21-C28	9625838.000	80.000	13.685	17.353	120322.975	125602.733	4.204
Aliphatic C28-C40	11438292.000	120.000	17.354	22.318	95319.100	101676.216	6.252
Aliphatic EPH	39961995.000	360.000	3.232	22.318	111005.542	116534.481	4.744