

Continuing Calibration Report for SequenceID : FF090325AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016313.D

Aromatic C10-C12	5203238.000	40.000	4.526	6.402	130080.950	133194.173	2.337
Aromatic C12-C16	7952856.000	60.000	6.403	9.106	132547.600	140138.141	5.416
Aromatic C16-C21	10388380.000	80.000	9.107	13.424	129854.750	139739.000	7.073
Aromatic C21-C36	21489318.000	180.000	13.425	18.875	119385.100	128435.755	7.047
Aromatic EPH	45033792.000	360.000	4.526	18.875	125093.867	133426.698	6.245