

Continuing Calibration Report for SequenceID : FF090325AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016323.D

Aromatic C10-C12	5357879.000	40.000	4.526	6.402	133946.975	133194.173	-0.565
Aromatic C12-C16	8176646.000	60.000	6.403	9.106	136277.433	140138.141	2.755
Aromatic C16-C21	10632539.000	80.000	9.107	13.424	132906.738	139739.000	4.889
Aromatic C21-C36	22117620.000	180.000	13.425	18.875	122875.667	128435.755	4.329
Aromatic EPH	46284684.000	360.000	4.526	18.875	128568.567	133426.698	3.641