

Continuing Calibration Report for SequenceID : FF090825AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FF016351.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 7222277.000  | 60.000  | 3.223  | 6.863  | 120371.283 | 118465.299 | -1.609 |
| Aliphatic C12-C16 | 4954612.000  | 40.000  | 8.597  | 10.307 | 123865.300 | 122693.653 | -0.955 |
| Aliphatic C16-C21 | 7775362.000  | 60.000  | 11.653 | 13.679 | 129589.367 | 128123.078 | -1.144 |
| Aliphatic C21-C28 | 10304402.000 | 80.000  | 14.166 | 17.348 | 128805.025 | 125602.733 | -2.550 |
| Aliphatic C28-C40 | 11673282.000 | 120.000 | 18.137 | 22.308 | 97277.350  | 101676.216 | 4.326  |
| Aliphatic EPH     | 41929935.000 | 360.000 | 3.223  | 22.308 | 116472.042 | 116534.481 | 0.054  |