

Continuing Calibration Report for SequenceID : FF090925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016354.D

Aliphatic C9-C12	7507854.000	60.000	3.229	6.866	125130.900	118465.299	-5.627
Aliphatic C12-C16	5176327.000	40.000	6.867	10.308	129408.175	122693.653	-5.473
Aliphatic C16-C21	8137898.000	60.000	10.309	13.678	135631.633	128123.078	-5.860
Aliphatic C21-C28	10720048.000	80.000	13.679	17.347	134000.600	125602.733	-6.686
Aliphatic C28-C40	12132775.000	120.000	17.348	22.306	101106.458	101676.216	0.560
Aliphatic EPH	43674902.000	360.000	3.229	22.306	121319.172	116534.481	-4.106