

Continuing Calibration Report for SequenceID : FF090925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016358.D

Aliphatic C9-C12	7631091.000	60.000	3.229	6.866	127184.850	118465.299	-7.360
Aliphatic C12-C16	5240603.000	40.000	6.867	10.308	131015.075	122693.653	-6.782
Aliphatic C16-C21	8262707.000	60.000	10.309	13.678	137711.783	128123.078	-7.484
Aliphatic C21-C28	10899064.000	80.000	13.679	17.347	136238.300	125602.733	-8.468
Aliphatic C28-C40	12382491.000	120.000	17.348	22.306	103187.425	101676.216	-1.486
Aliphatic EPH	44415956.000	360.000	3.229	22.306	123377.656	116534.481	-5.872