

Continuing Calibration Report for SequenceID : FG090925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016572.D

Aliphatic EPH	38593264.000	360.000	3.284	22.511	107203.511	106882.000	-0.301
Aliphatic C9-C12	6253765.000	60.000	3.284	6.929	104229.417	114749.036	9.168
Aliphatic C12-C16	4296634.000	40.000	6.930	10.389	107415.850	117373.513	8.484
Aliphatic C16-C21	6574143.000	60.000	10.390	13.779	109569.050	117474.265	6.729
Aliphatic C21-C28	9838584.000	80.000	13.780	17.466	122982.300	111788.212	-10.014
Aliphatic C28-C40	11630138.000	120.000	17.467	22.511	96917.817	90884.371	-6.639