

Continuing Calibration Report for SequenceID : FG090925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG016582.D

Aliphatic EPH	40679337.000	360.000	3.284	22.511	112998.158	106882.000	-5.722
Aliphatic C9-C12	7159095.000	60.000	3.284	6.929	119318.250	114749.036	-3.982
Aliphatic C12-C16	4925389.000	40.000	6.930	10.389	123134.725	117373.513	-4.908
Aliphatic C16-C21	7495240.000	60.000	10.390	13.779	124920.667	117474.265	-6.339
Aliphatic C21-C28	9602086.000	80.000	13.780	17.466	120026.075	111788.212	-7.369
Aliphatic C28-C40	11497527.000	120.000	17.467	22.511	95812.725	90884.371	-5.423