

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/11/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031697.D **Time Analyzed:** 10:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.584	9.483	9.683	795.690	712.500	11.7
2,4,5-TP(Silvex)	9.287	9.185	9.385	764.180	712.500	7.3
2,4-D	8.416	8.314	8.514	829.850	705.000	17.7
2,4-DB	10.160	10.058	10.258	813.640	712.500	14.2
2,4-DCAA	7.287	7.185	7.385	798.120	750.000	6.4
DICAMBA	7.471	7.369	7.569	753.080	705.000	6.8
DICHLORPROP	8.179	8.077	8.277	740.900	705.000	5.1
Dinoseb	11.356	11.254	11.454	829.460	705.000	17.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/11/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031697.D **Time Analyzed:** 10:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.323	10.221	10.421	732.550	712.500	2.8
2,4,5-TP(Silvex)	9.894	9.793	9.993	733.690	712.500	3.0
2,4-D	9.001	8.900	9.100	715.240	705.000	1.5
2,4-DB	10.891	10.790	10.990	691.840	712.500	-2.9
2,4-DCAA	7.744	7.643	7.843	755.620	750.000	0.7
DICAMBA	7.940	7.839	8.039	721.540	705.000	2.3
DICHLORPROP	8.659	8.559	8.759	733.760	705.000	4.1
Dinoseb	11.267	11.166	11.366	836.560	705.000	18.7