

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/11/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031705.D **Time Analyzed:** 18:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.585	9.483	9.683	791.670	712.500	11.1
2,4,5-TP(Silvex)	9.288	9.185	9.385	762.010	712.500	6.9
2,4-D	8.416	8.314	8.514	755.300	705.000	7.1
2,4-DB	10.160	10.058	10.258	790.870	712.500	11.0
2,4-DCAA	7.287	7.185	7.385	801.560	750.000	6.9
DICAMBA	7.471	7.369	7.569	756.190	705.000	7.3
DICHLORPROP	8.178	8.077	8.277	739.800	705.000	4.9
Dinoseb	11.357	11.254	11.454	823.460	705.000	16.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/11/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031705.D **Time Analyzed:** 18:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.324	10.221	10.421	734.750	712.500	3.1
2,4,5-TP(Silvex)	9.895	9.793	9.993	742.490	712.500	4.2
2,4-D	9.003	8.900	9.100	709.910	705.000	0.7
2,4-DB	10.892	10.790	10.990	705.700	712.500	-1.0
2,4-DCAA	7.744	7.643	7.843	756.270	750.000	0.8
DICAMBA	7.940	7.839	8.039	733.060	705.000	4.0
DICHLORPROP	8.660	8.559	8.759	742.610	705.000	5.3
Dinoseb	11.267	11.166	11.366	831.050	705.000	17.9