

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/12/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031717.D **Time Analyzed:** 10:18

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.584	9.483	9.683	793.450	712.500	11.4
2,4,5-TP(Silvex)	9.286	9.185	9.385	765.270	712.500	7.4
2,4-D	8.414	8.314	8.514	740.420	705.000	5.0
2,4-DB	10.158	10.058	10.258	713.130	712.500	0.1
2,4-DCAA	7.286	7.185	7.385	795.330	750.000	6.0
DICAMBA	7.470	7.369	7.569	755.780	705.000	7.2
DICHLORPROP	8.178	8.077	8.277	738.780	705.000	4.8
Dinoseb	11.354	11.254	11.454	758.100	705.000	7.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/12/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031717.D **Time Analyzed:** 10:18

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.322	10.221	10.421	730.730	712.500	2.6
2,4,5-TP(Silvex)	9.893	9.793	9.993	736.160	712.500	3.3
2,4-D	9.001	8.900	9.100	706.550	705.000	0.2
2,4-DB	10.890	10.790	10.990	691.940	712.500	-2.9
2,4-DCAA	7.743	7.643	7.843	748.190	750.000	-0.2
DICAMBA	7.939	7.839	8.039	722.130	705.000	2.4
DICHLORPROP	8.659	8.559	8.759	734.210	705.000	4.1
Dinoseb	11.265	11.166	11.366	835.280	705.000	18.5