

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ALLI03  
**Lab Code:** ACE **SDG NO.:** Q2893  
**GC Column:** RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

**Client Sample No.:** CCAL03 **Date Analyzed:** 09/16/2025  
**Lab Sample No.:** HSTDCCC750 **Data File :** PS031750.D **Time Analyzed:** 10:15

| COMPOUND         | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|--------|-----------|--------|--------------------|-------------------|------|
|                  |        | FROM      | TO     |                    |                   |      |
| 2,4,5-T          | 9.582  | 9.483     | 9.683  | 770.690            | 712.500           | 8.2  |
| 2,4,5-TP(Silvex) | 9.284  | 9.185     | 9.385  | 732.480            | 712.500           | 2.8  |
| 2,4-D            | 8.414  | 8.314     | 8.514  | 791.980            | 705.000           | 12.3 |
| 2,4-DB           | 10.156 | 10.058    | 10.258 | 727.350            | 712.500           | 2.1  |
| 2,4-DCAA         | 7.285  | 7.185     | 7.385  | 763.690            | 750.000           | 1.8  |
| DICAMBA          | 7.469  | 7.369     | 7.569  | 722.100            | 705.000           | 2.4  |
| DICHLORPROP      | 8.177  | 8.077     | 8.277  | 707.120            | 705.000           | 0.3  |
| Dinoseb          | 11.354 | 11.254    | 11.454 | 738.870            | 705.000           | 4.8  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ALLI03  
**Lab Code:** ACE **SDG NO.:** Q2893  
**GC Column:** RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

**Client Sample No.:** CCAL03 **Date Analyzed:** 09/16/2025  
**Lab Sample No.:** HSTDCCC750 **Data File :** PS031750.D **Time Analyzed:** 10:15

| COMPOUND         | RT     | RT WINDOW<br>FROM TO |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|------------------|--------|----------------------|--------|--------------------|-------------------|------|
| 2,4,5-T          | 10.322 | 10.221               | 10.421 | 696.810            | 712.500           | -2.2 |
| 2,4,5-TP(Silvex) | 9.893  | 9.793                | 9.993  | 703.710            | 712.500           | -1.2 |
| 2,4-D            | 9.001  | 8.900                | 9.100  | 683.260            | 705.000           | -3.1 |
| 2,4-DB           | 10.889 | 10.790               | 10.990 | 644.940            | 712.500           | -9.5 |
| 2,4-DCAA         | 7.743  | 7.643                | 7.843  | 722.840            | 750.000           | -3.6 |
| DICAMBA          | 7.939  | 7.839                | 8.039  | 695.280            | 705.000           | -1.4 |
| DICHLORPROP      | 8.658  | 8.559                | 8.759  | 704.060            | 705.000           | -0.1 |
| Dinoseb          | 11.265 | 11.166               | 11.366 | 722.910            | 705.000           | 2.5  |