

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: ALLI03

Lab Code: ACE SDG NO.: Q2893

Instrument ID: ECD_P Date(s) Analyzed: 08/19/2025 08/19/2025

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.85	4.75	4.95	17592800
		2	4.94	4.84	5.04	13357200
		3	5.02	4.92	5.12	37912800
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	5.02	4.92	5.12	30374000
		2	5.54	5.44	5.64	15758200
		3	5.83	5.73	5.93	30543600
		4	5.99	5.89	6.09	17269300
		5	6.08	5.98	6.18	12981000
Aroclor-1262	500	1	8.24	8.14	8.34	77016000
		2	8.57	8.47	8.67	132381000
		3	8.89	8.79	8.99	97212400
		4	8.98	8.88	9.08	75330400
		5	9.65	9.55	9.75	60956200

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Lab Name: Alliance Contract: ALLI03

Lab Code: ACE SDG NO.: Q2893

Instrument ID: ECD_P Date(s) Analyzed: 08/19/2025 08/19/2025

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Aroclor-1221	500	1	4.01	3.91	4.11	61867400
		2	4.10	4.00	4.20	45212000
		3	4.17	4.07	4.27	164294000
		4	0.00			0
		5	0.00			0
Aroclor-1232	500	1	4.17	4.07	4.27	128095000
		2	4.90	4.80	5.00	231794000
		3	5.08	4.98	5.18	60059800
		4	5.16	5.06	5.26	73707400
		5	5.33	5.23	5.43	62083800
Aroclor-1262	500	1	6.92	6.82	7.02	707260000
		2	7.18	7.08	7.28	520714000
		3	7.70	7.60	7.80	466296000
		4	7.76	7.66	7.86	821860000
		5	8.26	8.16	8.36	358310000