

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q2893

Continuing Calib Date: 09/04/2025

Initial Calibration Date(s): 08/18/2025

08/18/2025

Continuing Calib Time: 09:49

Initial Calibration Time(s): 09:53

18:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.55	5.55	5.45	5.65	0.00
Aroclor-1254-2 (2)	5.70	5.70	5.60	5.80	0.00
Aroclor-1254-3 (3)	6.10	6.11	6.01	6.21	0.01
Aroclor-1254-4 (4)	6.33	6.34	6.24	6.44	0.01
Aroclor-1254-5 (5)	6.75	6.75	6.65	6.85	0.00
Tetrachloro-m-xylene	3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl	8.68	8.68	8.58	8.78	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q2893

Continuing Calib Date: 09/04/2025

Initial Calibration Date(s): 08/18/2025

08/18/2025

Continuing Calib Time: 09:49

Initial Calibration Time(s): 09:53

18:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.52	5.53	5.43	5.63	0.01
Aroclor-1254-2	(2)	5.67	5.67	5.57	5.77	0.00
Aroclor-1254-3	(3)	6.07	6.07	5.97	6.17	0.00
Aroclor-1254-4	(4)	6.30	6.30	6.20	6.40	0.00
Aroclor-1254-5	(5)	6.71	6.72	6.62	6.82	0.01
Tetrachloro-m-xylene		3.65	3.66	3.56	3.76	0.01
Decachlorobiphenyl		8.62	8.63	8.53	8.73	0.01