

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q2893

Continuing Calib Date: 09/08/2025

Initial Calibration Date(s): 08/19/2025

08/19/2025

Continuing Calib Time: 12:11

Initial Calibration Time(s): 10:02

17:36

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|-------|------------|
| Aroclor-1242-1 (1)   | 5.81       | 5.81      | 5.71                   | 5.91  | 0.00       |
| Aroclor-1242-2 (2)   | 5.83       | 5.83      | 5.73                   | 5.93  | 0.00       |
| Aroclor-1242-3 (3)   | 5.90       | 5.89      | 5.79                   | 5.99  | -0.01      |
| Aroclor-1242-4 (4)   | 5.99       | 5.99      | 5.89                   | 6.09  | 0.00       |
| Aroclor-1242-5 (5)   | 6.72       | 6.72      | 6.62                   | 6.82  | 0.00       |
| Tetrachloro-m-xylene | 4.66       | 4.66      | 4.56                   | 4.76  | 0.00       |
| Decachlorobiphenyl   | 10.44      | 10.43     | 10.33                  | 10.53 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q2893

Continuing Calib Date: 09/08/2025

Initial Calibration Date(s): 08/19/2025

08/19/2025

Continuing Calib Time: 12:11

Initial Calibration Time(s): 10:02

17:36

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1242-1 (1)   | 4.90       | 4.90      | 4.80      | 5.00 | 0.00       |
| Aroclor-1242-2 (2)   | 4.96       | 4.96      | 4.86      | 5.06 | 0.01       |
| Aroclor-1242-3 (3)   | 5.08       | 5.08      | 4.98      | 5.18 | 0.01       |
| Aroclor-1242-4 (4)   | 5.16       | 5.16      | 5.06      | 5.26 | 0.00       |
| Aroclor-1242-5 (5)   | 5.68       | 5.68      | 5.58      | 5.78 | 0.00       |
| Tetrachloro-m-xylene | 3.80       | 3.80      | 3.70      | 3.90 | 0.00       |
| Decachlorobiphenyl   | 8.81       | 8.81      | 8.71      | 8.91 | 0.00       |