

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q2893

Continuing Calib Date: 09/10/2025

Initial Calibration Date(s): 09/10/2025 09/10/2025

Continuing Calib Time: 12:39

Initial Calibration Time(s): 09:25 12:05

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.49       | 4.48      | 4.38      | 4.58 | -0.01      |
| Aroclor-1016-2       | (2) | 4.51       | 4.50      | 4.40      | 4.60 | -0.01      |
| Aroclor-1016-3       | (3) | 4.57       | 4.56      | 4.46      | 4.66 | -0.01      |
| Aroclor-1016-4       | (4) | 4.66       | 4.65      | 4.55      | 4.75 | -0.01      |
| Aroclor-1016-5       | (5) | 4.94       | 4.93      | 4.83      | 5.03 | -0.01      |
| Aroclor-1260-1       | (1) | 6.03       | 6.02      | 5.92      | 6.12 | -0.01      |
| Aroclor-1260-2       | (2) | 6.29       | 6.28      | 6.18      | 6.38 | -0.01      |
| Aroclor-1260-3       | (3) | 6.64       | 6.63      | 6.53      | 6.73 | -0.01      |
| Aroclor-1260-4       | (4) | 6.85       | 6.84      | 6.74      | 6.94 | -0.01      |
| Aroclor-1260-5       | (5) | 7.16       | 7.15      | 7.05      | 7.25 | -0.01      |
| Tetrachloro-m-xylene |     | 3.41       | 3.40      | 3.30      | 3.50 | -0.01      |
| Decachlorobiphenyl   |     | 8.51       | 8.50      | 8.40      | 8.60 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q2893

Continuing Calib Date: 09/10/2025

Initial Calibration Date(s): 09/10/2025 09/10/2025

Continuing Calib Time: 12:39

Initial Calibration Time(s): 09:25 12:05

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 3.78       | 3.78      | 3.68      | 3.88 | 0.00       |
| Aroclor-1016-2       | (2) | 3.79       | 3.79      | 3.69      | 3.89 | 0.00       |
| Aroclor-1016-3       | (3) | 3.95       | 3.95      | 3.85      | 4.05 | 0.00       |
| Aroclor-1016-4       | (4) | 4.00       | 4.00      | 3.90      | 4.10 | 0.00       |
| Aroclor-1016-5       | (5) | 4.20       | 4.20      | 4.10      | 4.30 | 0.00       |
| Aroclor-1260-1       | (1) | 5.19       | 5.19      | 5.09      | 5.29 | 0.00       |
| Aroclor-1260-2       | (2) | 5.38       | 5.38      | 5.28      | 5.48 | 0.00       |
| Aroclor-1260-3       | (3) | 5.52       | 5.52      | 5.42      | 5.62 | 0.00       |
| Aroclor-1260-4       | (4) | 5.98       | 5.98      | 5.88      | 6.08 | 0.00       |
| Aroclor-1260-5       | (5) | 6.23       | 6.23      | 6.13      | 6.33 | 0.00       |
| Tetrachloro-m-xylene |     | 2.78       | 2.78      | 2.68      | 2.88 | 0.00       |
| Decachlorobiphenyl   |     | 7.54       | 7.54      | 7.44      | 7.64 | 0.00       |