

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113449.D **Time Analyzed:** 08:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.743	4.648	4.848	583.050	500.000	16.6
Aroclor-1016-2	4.761	4.666	4.866	596.580	500.000	19.3
Aroclor-1016-3	4.818	4.724	4.924	592.410	500.000	18.5
Aroclor-1016-4	4.937	4.843	5.043	543.880	500.000	8.8
Aroclor-1016-5	5.194	5.099	5.299	570.150	500.000	14.0
Aroclor-1260-1	6.230	6.136	6.336	572.490	500.000	14.5
Aroclor-1260-2	6.420	6.326	6.526	564.240	500.000	12.8
Aroclor-1260-3	6.785	6.691	6.891	566.210	500.000	13.2
Aroclor-1260-4	7.044	6.951	7.151	557.940	500.000	11.6
Aroclor-1260-5	7.288	7.194	7.394	554.350	500.000	10.9
Decachlorobiphenyl	8.675	8.582	8.782	50.440	50.000	0.9
Tetrachloro-m-xylene	3.659	3.563	3.763	59.220	50.000	18.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113449.D **Time Analyzed:** 08:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.723	4.629	4.829	541.170	500.000	8.2
Aroclor-1016-2	4.742	4.647	4.847	540.550	500.000	8.1
Aroclor-1016-3	4.917	4.822	5.022	514.040	500.000	2.8
Aroclor-1016-4	4.958	4.864	5.064	502.980	500.000	0.6
Aroclor-1016-5	5.171	5.076	5.276	540.890	500.000	8.2
Aroclor-1260-1	6.198	6.104	6.304	500.230	500.000	0.0
Aroclor-1260-2	6.386	6.292	6.492	485.330	500.000	-2.9
Aroclor-1260-3	6.538	6.443	6.643	462.210	500.000	-7.6
Aroclor-1260-4	7.007	6.912	7.112	473.210	500.000	-5.4
Aroclor-1260-5	7.249	7.155	7.355	462.350	500.000	-7.5
Decachlorobiphenyl	8.620	8.528	8.728	53.820	50.000	7.6
Tetrachloro-m-xylene	3.650	3.555	3.755	55.530	50.000	11.1