

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113463.D **Time Analyzed:** 13:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.743	4.648	4.848	553.950	500.000	10.8
Aroclor-1016-2	4.762	4.666	4.866	565.510	500.000	13.1
Aroclor-1016-3	4.819	4.724	4.924	561.300	500.000	12.3
Aroclor-1016-4	4.937	4.843	5.043	499.850	500.000	0.0
Aroclor-1016-5	5.194	5.099	5.299	557.710	500.000	11.5
Aroclor-1260-1	6.231	6.136	6.336	588.590	500.000	17.7
Aroclor-1260-2	6.420	6.326	6.526	559.940	500.000	12.0
Aroclor-1260-3	6.786	6.691	6.891	548.450	500.000	9.7
Aroclor-1260-4	7.045	6.951	7.151	549.200	500.000	9.8
Aroclor-1260-5	7.288	7.194	7.394	548.620	500.000	9.7
Decachlorobiphenyl	8.675	8.582	8.782	51.490	50.000	3.0
Tetrachloro-m-xylene	3.659	3.563	3.763	56.340	50.000	12.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113463.D **Time Analyzed:** 13:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.723	4.629	4.829	519.310	500.000	3.9
Aroclor-1016-2	4.741	4.647	4.847	520.160	500.000	4.0
Aroclor-1016-3	4.916	4.822	5.022	491.590	500.000	-1.7
Aroclor-1016-4	4.959	4.864	5.064	480.660	500.000	-3.9
Aroclor-1016-5	5.170	5.076	5.276	528.420	500.000	5.7
Aroclor-1260-1	6.198	6.104	6.304	480.330	500.000	-3.9
Aroclor-1260-2	6.386	6.292	6.492	467.180	500.000	-6.6
Aroclor-1260-3	6.537	6.443	6.643	442.500	500.000	-11.5
Aroclor-1260-4	7.007	6.912	7.112	452.620	500.000	-9.5
Aroclor-1260-5	7.249	7.155	7.355	445.070	500.000	-11.0
Decachlorobiphenyl	8.620	8.528	8.728	54.150	50.000	8.3
Tetrachloro-m-xylene	3.650	3.555	3.755	53.360	50.000	6.7