

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113523.D **Time Analyzed:** 20:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.744	4.648	4.848	545.580	500.000	9.1
Aroclor-1016-2	4.762	4.666	4.866	554.040	500.000	10.8
Aroclor-1016-3	4.819	4.724	4.924	556.660	500.000	11.3
Aroclor-1016-4	4.938	4.843	5.043	510.110	500.000	2.0
Aroclor-1016-5	5.195	5.099	5.299	581.150	500.000	16.2
Aroclor-1260-1	6.230	6.136	6.336	491.370	500.000	-1.7
Aroclor-1260-2	6.420	6.326	6.526	480.520	500.000	-3.9
Aroclor-1260-3	6.785	6.691	6.891	506.650	500.000	1.3
Aroclor-1260-4	7.044	6.951	7.151	553.810	500.000	10.8
Aroclor-1260-5	7.288	7.194	7.394	475.090	500.000	-5.0
Decachlorobiphenyl	8.674	8.582	8.782	41.190	50.000	-17.6
Tetrachloro-m-xylene	3.660	3.563	3.763	54.060	50.000	8.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113523.D **Time Analyzed:** 20:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.724	4.629	4.829	517.300	500.000	3.5
Aroclor-1016-2	4.742	4.647	4.847	520.370	500.000	4.1
Aroclor-1016-3	4.917	4.822	5.022	499.100	500.000	-0.2
Aroclor-1016-4	4.960	4.864	5.064	484.780	500.000	-3.0
Aroclor-1016-5	5.172	5.076	5.276	545.070	500.000	9.0
Aroclor-1260-1	6.198	6.104	6.304	479.800	500.000	-4.0
Aroclor-1260-2	6.386	6.292	6.492	455.010	500.000	-9.0
Aroclor-1260-3	6.537	6.443	6.643	418.420	500.000	-16.3
Aroclor-1260-4	7.007	6.912	7.112	426.820	500.000	-14.6
Aroclor-1260-5	7.249	7.155	7.355	413.560	500.000	-17.3
Decachlorobiphenyl	8.619	8.528	8.728	48.040	50.000	-3.9
Tetrachloro-m-xylene	3.652	3.555	3.755	52.210	50.000	4.4