

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113647.D **Time Analyzed:** 09:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.740	4.648	4.848	562.590	500.000	12.5
Aroclor-1016-2	4.758	4.666	4.866	579.800	500.000	16.0
Aroclor-1016-3	4.815	4.724	4.924	571.560	500.000	14.3
Aroclor-1016-4	4.934	4.843	5.043	536.850	500.000	7.4
Aroclor-1016-5	5.191	5.099	5.299	552.420	500.000	10.5
Aroclor-1260-1	6.226	6.136	6.336	571.890	500.000	14.4
Aroclor-1260-2	6.416	6.326	6.526	569.560	500.000	13.9
Aroclor-1260-3	6.781	6.691	6.891	582.260	500.000	16.5
Aroclor-1260-4	7.039	6.951	7.151	579.910	500.000	16.0
Aroclor-1260-5	7.282	7.194	7.394	574.420	500.000	14.9
Decachlorobiphenyl	8.667	8.582	8.782	56.150	50.000	12.3
Tetrachloro-m-xylene	3.657	3.563	3.763	59.110	50.000	18.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113647.D **Time Analyzed:** 09:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.720	4.629	4.829	508.200	500.000	1.6
Aroclor-1016-2	4.738	4.647	4.847	527.970	500.000	5.6
Aroclor-1016-3	4.913	4.822	5.022	517.950	500.000	3.6
Aroclor-1016-4	4.954	4.864	5.064	510.590	500.000	2.1
Aroclor-1016-5	5.167	5.076	5.276	534.260	500.000	6.9
Aroclor-1260-1	6.193	6.104	6.304	490.060	500.000	-2.0
Aroclor-1260-2	6.381	6.292	6.492	464.420	500.000	-7.1
Aroclor-1260-3	6.532	6.443	6.643	446.640	500.000	-10.7
Aroclor-1260-4	7.000	6.912	7.112	449.000	500.000	-10.2
Aroclor-1260-5	7.244	7.155	7.355	441.570	500.000	-11.7
Decachlorobiphenyl	8.612	8.528	8.728	56.410	50.000	12.8
Tetrachloro-m-xylene	3.648	3.555	3.755	55.500	50.000	11.0