

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074839.D **Time Analyzed:** 08:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.808	5.706	5.906	491.750	500.000	-1.7
Aroclor-1016-2	5.830	5.727	5.927	516.290	500.000	3.3
Aroclor-1016-3	5.893	5.790	5.990	467.140	500.000	-6.6
Aroclor-1016-4	5.990	5.888	6.088	492.430	500.000	-1.5
Aroclor-1016-5	6.282	6.180	6.380	511.520	500.000	2.3
Aroclor-1260-1	7.399	7.297	7.497	474.600	500.000	-5.1
Aroclor-1260-2	7.652	7.550	7.750	465.770	500.000	-6.8
Aroclor-1260-3	8.010	7.908	8.108	491.170	500.000	-1.8
Aroclor-1260-4	8.238	8.136	8.336	471.410	500.000	-5.7
Aroclor-1260-5	8.564	8.462	8.662	450.750	500.000	-9.9
Decachlorobiphenyl	10.435	10.333	10.533	46.120	50.000	-7.8
Tetrachloro-m-xylene	4.656	4.554	4.754	51.410	50.000	2.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074839.D **Time Analyzed:** 08:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.896	4.799	4.999	531.740	500.000	6.3
Aroclor-1016-2	4.954	4.855	5.055	551.080	500.000	10.2
Aroclor-1016-3	5.074	4.976	5.176	519.170	500.000	3.8
Aroclor-1016-4	5.114	5.017	5.217	526.530	500.000	5.3
Aroclor-1016-5	5.329	5.231	5.431	481.360	500.000	-3.7
Aroclor-1260-1	6.359	6.261	6.461	478.750	500.000	-4.3
Aroclor-1260-2	6.545	6.448	6.648	455.630	500.000	-8.9
Aroclor-1260-3	6.699	6.602	6.802	429.500	500.000	-14.1
Aroclor-1260-4	7.169	7.072	7.272	445.290	500.000	-10.9
Aroclor-1260-5	7.408	7.311	7.511	434.960	500.000	-13.0
Decachlorobiphenyl	8.809	8.715	8.915	40.810	50.000	-18.4
Tetrachloro-m-xylene	3.798	3.699	3.899	59.720	50.000	19.4