

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074850.D **Time Analyzed:** 12:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.814	5.706	5.906	482.470	500.000	-3.5
Aroclor-1016-2	5.834	5.727	5.927	475.470	500.000	-4.9
Aroclor-1016-3	5.898	5.790	5.990	457.000	500.000	-8.6
Aroclor-1016-4	5.995	5.888	6.088	483.390	500.000	-3.3
Aroclor-1016-5	6.288	6.180	6.380	500.710	500.000	0.1
Aroclor-1260-1	7.406	7.297	7.497	465.030	500.000	-7.0
Aroclor-1260-2	7.659	7.550	7.750	459.240	500.000	-8.2
Aroclor-1260-3	8.017	7.908	8.108	486.890	500.000	-2.6
Aroclor-1260-4	8.245	8.136	8.336	465.360	500.000	-6.9
Aroclor-1260-5	8.572	8.462	8.662	450.400	500.000	-9.9
Decachlorobiphenyl	10.445	10.333	10.533	53.560	50.000	7.1
Tetrachloro-m-xylene	4.662	4.554	4.754	49.680	50.000	-0.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074850.D **Time Analyzed:** 12:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.902	4.799	4.999	508.270	500.000	1.7
Aroclor-1016-2	4.959	4.855	5.055	520.170	500.000	4.0
Aroclor-1016-3	5.080	4.976	5.176	498.300	500.000	-0.3
Aroclor-1016-4	5.120	5.017	5.217	498.140	500.000	-0.4
Aroclor-1016-5	5.334	5.231	5.431	463.680	500.000	-7.3
Aroclor-1260-1	6.365	6.261	6.461	463.900	500.000	-7.2
Aroclor-1260-2	6.551	6.448	6.648	452.830	500.000	-9.4
Aroclor-1260-3	6.706	6.602	6.802	410.570	500.000	-17.9
Aroclor-1260-4	7.174	7.072	7.272	439.760	500.000	-12.0
Aroclor-1260-5	7.414	7.311	7.511	429.600	500.000	-14.1
Decachlorobiphenyl	8.816	8.715	8.915	49.040	50.000	-1.9
Tetrachloro-m-xylene	3.804	3.699	3.899	54.310	50.000	8.6