

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074861.D **Time Analyzed:** 15:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.814	5.706	5.906	483.860	500.000	-3.2
Aroclor-1016-2	5.835	5.727	5.927	469.130	500.000	-6.2
Aroclor-1016-3	5.899	5.790	5.990	454.250	500.000	-9.2
Aroclor-1016-4	5.996	5.888	6.088	483.730	500.000	-3.3
Aroclor-1016-5	6.288	6.180	6.380	502.220	500.000	0.4
Aroclor-1260-1	7.406	7.297	7.497	467.330	500.000	-6.5
Aroclor-1260-2	7.659	7.550	7.750	464.400	500.000	-7.1
Aroclor-1260-3	8.018	7.908	8.108	494.650	500.000	-1.1
Aroclor-1260-4	8.246	8.136	8.336	471.040	500.000	-5.8
Aroclor-1260-5	8.573	8.462	8.662	452.270	500.000	-9.5
Decachlorobiphenyl	10.447	10.333	10.533	54.000	50.000	8.0
Tetrachloro-m-xylene	4.663	4.554	4.754	50.030	50.000	0.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074861.D **Time Analyzed:** 15:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.901	4.799	4.999	508.330	500.000	1.7
Aroclor-1016-2	4.958	4.855	5.055	511.910	500.000	2.4
Aroclor-1016-3	5.079	4.976	5.176	502.100	500.000	0.4
Aroclor-1016-4	5.120	5.017	5.217	486.240	500.000	-2.8
Aroclor-1016-5	5.334	5.231	5.431	440.330	500.000	-11.9
Aroclor-1260-1	6.364	6.261	6.461	456.420	500.000	-8.7
Aroclor-1260-2	6.550	6.448	6.648	447.120	500.000	-10.6
Aroclor-1260-3	6.704	6.602	6.802	409.900	500.000	-18.0
Aroclor-1260-4	7.173	7.072	7.272	447.300	500.000	-10.5
Aroclor-1260-5	7.413	7.311	7.511	430.300	500.000	-13.9
Decachlorobiphenyl	8.816	8.715	8.915	50.080	50.000	0.2
Tetrachloro-m-xylene	3.802	3.699	3.899	56.690	50.000	13.4