

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074906.D **Time Analyzed:** 15:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.811	5.706	5.906	486.510	500.000	-2.7
Aroclor-1016-2	5.833	5.727	5.927	506.700	500.000	1.3
Aroclor-1016-3	5.896	5.790	5.990	462.200	500.000	-7.6
Aroclor-1016-4	5.993	5.888	6.088	485.780	500.000	-2.8
Aroclor-1016-5	6.286	6.180	6.380	502.140	500.000	0.4
Aroclor-1260-1	7.403	7.297	7.497	475.740	500.000	-4.9
Aroclor-1260-2	7.655	7.550	7.750	466.810	500.000	-6.6
Aroclor-1260-3	8.014	7.908	8.108	493.970	500.000	-1.2
Aroclor-1260-4	8.242	8.136	8.336	474.090	500.000	-5.2
Aroclor-1260-5	8.568	8.462	8.662	465.510	500.000	-6.9
Decachlorobiphenyl	10.440	10.333	10.533	47.250	50.000	-5.5
Tetrachloro-m-xylene	4.660	4.554	4.754	50.390	50.000	0.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074906.D **Time Analyzed:** 15:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.899	4.799	4.999	525.300	500.000	5.1
Aroclor-1016-2	4.957	4.855	5.055	522.200	500.000	4.4
Aroclor-1016-3	5.077	4.976	5.176	496.610	500.000	-0.7
Aroclor-1016-4	5.118	5.017	5.217	509.060	500.000	1.8
Aroclor-1016-5	5.332	5.231	5.431	464.210	500.000	-7.2
Aroclor-1260-1	6.361	6.261	6.461	479.420	500.000	-4.1
Aroclor-1260-2	6.548	6.448	6.648	471.010	500.000	-5.8
Aroclor-1260-3	6.701	6.602	6.802	434.280	500.000	-13.1
Aroclor-1260-4	7.170	7.072	7.272	479.070	500.000	-4.2
Aroclor-1260-5	7.410	7.311	7.511	462.140	500.000	-7.6
Decachlorobiphenyl	8.811	8.715	8.915	45.420	50.000	-9.2
Tetrachloro-m-xylene	3.800	3.699	3.899	58.340	50.000	16.7