

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074942.D **Time Analyzed:** 09:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.801	5.706	5.906	491.120	500.000	-1.8
Aroclor-1016-2	5.823	5.727	5.927	510.350	500.000	2.1
Aroclor-1016-3	5.886	5.790	5.990	457.170	500.000	-8.6
Aroclor-1016-4	5.982	5.888	6.088	484.230	500.000	-3.2
Aroclor-1016-5	6.275	6.180	6.380	499.530	500.000	-0.1
Aroclor-1260-1	7.392	7.297	7.497	459.590	500.000	-8.1
Aroclor-1260-2	7.644	7.550	7.750	453.560	500.000	-9.3
Aroclor-1260-3	8.002	7.908	8.108	484.810	500.000	-3.0
Aroclor-1260-4	8.230	8.136	8.336	453.690	500.000	-9.3
Aroclor-1260-5	8.555	8.462	8.662	452.790	500.000	-9.4
Decachlorobiphenyl	10.421	10.333	10.533	45.760	50.000	-8.5
Tetrachloro-m-xylene	4.650	4.554	4.754	49.650	50.000	-0.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074942.D **Time Analyzed:** 09:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.892	4.799	4.999	494.830	500.000	-1.0
Aroclor-1016-2	4.949	4.855	5.055	520.270	500.000	4.1
Aroclor-1016-3	5.069	4.976	5.176	500.060	500.000	0.0
Aroclor-1016-4	5.110	5.017	5.217	491.110	500.000	-1.8
Aroclor-1016-5	5.324	5.231	5.431	453.320	500.000	-9.3
Aroclor-1260-1	6.353	6.261	6.461	459.120	500.000	-8.2
Aroclor-1260-2	6.539	6.448	6.648	445.070	500.000	-11.0
Aroclor-1260-3	6.693	6.602	6.802	419.000	500.000	-16.2
Aroclor-1260-4	7.162	7.072	7.272	434.650	500.000	-13.1
Aroclor-1260-5	7.401	7.311	7.511	439.280	500.000	-12.1
Decachlorobiphenyl	8.802	8.715	8.915	42.930	50.000	-14.1
Tetrachloro-m-xylene	3.794	3.699	3.899	55.610	50.000	11.2