

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074949.D **Time Analyzed:** 11:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.805	5.706	5.906	486.060	500.000	-2.8
Aroclor-1016-2	5.827	5.727	5.927	509.480	500.000	1.9
Aroclor-1016-3	5.889	5.790	5.990	453.190	500.000	-9.4
Aroclor-1016-4	5.986	5.888	6.088	485.000	500.000	-3.0
Aroclor-1016-5	6.279	6.180	6.380	499.310	500.000	-0.1
Aroclor-1260-1	7.396	7.297	7.497	469.110	500.000	-6.2
Aroclor-1260-2	7.649	7.550	7.750	463.930	500.000	-7.2
Aroclor-1260-3	8.007	7.908	8.108	493.480	500.000	-1.3
Aroclor-1260-4	8.235	8.136	8.336	461.410	500.000	-7.7
Aroclor-1260-5	8.560	8.462	8.662	455.280	500.000	-8.9
Decachlorobiphenyl	10.429	10.333	10.533	46.870	50.000	-6.3
Tetrachloro-m-xylene	4.654	4.554	4.754	49.180	50.000	-1.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/08/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074949.D **Time Analyzed:** 11:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.895	4.799	4.999	478.730	500.000	-4.3
Aroclor-1016-2	4.952	4.855	5.055	499.640	500.000	-0.1
Aroclor-1016-3	5.072	4.976	5.176	496.040	500.000	-0.8
Aroclor-1016-4	5.114	5.017	5.217	475.910	500.000	-4.8
Aroclor-1016-5	5.327	5.231	5.431	436.740	500.000	-12.7
Aroclor-1260-1	6.356	6.261	6.461	457.050	500.000	-8.6
Aroclor-1260-2	6.542	6.448	6.648	446.600	500.000	-10.7
Aroclor-1260-3	6.696	6.602	6.802	414.260	500.000	-17.1
Aroclor-1260-4	7.166	7.072	7.272	434.920	500.000	-13.0
Aroclor-1260-5	7.405	7.311	7.511	432.890	500.000	-13.4
Decachlorobiphenyl	8.806	8.715	8.915	42.240	50.000	-15.5
Tetrachloro-m-xylene	3.797	3.699	3.899	54.710	50.000	9.4