

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/10/2025 09/10/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ070813.D **Time Analyzed:** 15:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.482	4.381	4.581	516.960	500.000	3.4
Aroclor-1016-2	4.501	4.400	4.600	504.830	500.000	1.0
Aroclor-1016-3	4.558	4.457	4.657	511.090	500.000	2.2
Aroclor-1016-4	4.648	4.548	4.748	507.670	500.000	1.5
Aroclor-1016-5	4.932	4.832	5.032	513.820	500.000	2.8
Aroclor-1260-1	6.021	5.921	6.121	502.300	500.000	0.5
Aroclor-1260-2	6.277	6.177	6.377	487.100	500.000	-2.6
Aroclor-1260-3	6.628	6.528	6.728	512.000	500.000	2.4
Aroclor-1260-4	6.841	6.742	6.942	508.660	500.000	1.7
Aroclor-1260-5	7.147	7.047	7.247	499.080	500.000	-0.2
Decachlorobiphenyl	8.498	8.399	8.599	50.780	50.000	1.6
Tetrachloro-m-xylene	3.401	3.302	3.502	50.730	50.000	1.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/10/2025 09/10/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ070813.D **Time Analyzed:** 15:41

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.779	3.678	3.878	467.680	500.000	-6.5
Aroclor-1016-2	3.794	3.694	3.894	478.840	500.000	-4.2
Aroclor-1016-3	3.954	3.854	4.054	469.550	500.000	-6.1
Aroclor-1016-4	4.003	3.903	4.103	470.840	500.000	-5.8
Aroclor-1016-5	4.198	4.097	4.297	480.850	500.000	-3.8
Aroclor-1260-1	5.190	5.090	5.290	481.770	500.000	-3.6
Aroclor-1260-2	5.380	5.280	5.480	478.060	500.000	-4.4
Aroclor-1260-3	5.521	5.420	5.620	478.210	500.000	-4.4
Aroclor-1260-4	5.981	5.880	6.080	478.020	500.000	-4.4
Aroclor-1260-5	6.225	6.126	6.326	466.600	500.000	-6.7
Decachlorobiphenyl	7.535	7.436	7.636	48.280	50.000	-3.4
Tetrachloro-m-xylene	2.779	2.678	2.878	49.710	50.000	-0.6