

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/10/2025 09/10/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ070824.D **Time Analyzed:** 18:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.481	4.381	4.581	518.530	500.000	3.7
Aroclor-1016-2	4.500	4.400	4.600	505.190	500.000	1.0
Aroclor-1016-3	4.557	4.457	4.657	511.570	500.000	2.3
Aroclor-1016-4	4.647	4.548	4.748	505.660	500.000	1.1
Aroclor-1016-5	4.931	4.832	5.032	509.810	500.000	2.0
Aroclor-1260-1	6.021	5.921	6.121	499.530	500.000	-0.1
Aroclor-1260-2	6.276	6.177	6.377	489.350	500.000	-2.1
Aroclor-1260-3	6.627	6.528	6.728	512.990	500.000	2.6
Aroclor-1260-4	6.842	6.742	6.942	505.160	500.000	1.0
Aroclor-1260-5	7.147	7.047	7.247	499.690	500.000	-0.1
Decachlorobiphenyl	8.497	8.399	8.599	50.420	50.000	0.8
Tetrachloro-m-xylene	3.401	3.302	3.502	50.970	50.000	1.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/10/2025 09/10/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PQ070824.D **Time Analyzed:** 18:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.779	3.678	3.878	470.070	500.000	-6.0
Aroclor-1016-2	3.794	3.694	3.894	471.260	500.000	-5.7
Aroclor-1016-3	3.954	3.854	4.054	467.280	500.000	-6.5
Aroclor-1016-4	4.003	3.903	4.103	470.420	500.000	-5.9
Aroclor-1016-5	4.198	4.097	4.297	476.220	500.000	-4.8
Aroclor-1260-1	5.190	5.090	5.290	479.510	500.000	-4.1
Aroclor-1260-2	5.380	5.280	5.480	474.640	500.000	-5.1
Aroclor-1260-3	5.521	5.420	5.620	476.770	500.000	-4.6
Aroclor-1260-4	5.981	5.880	6.080	475.290	500.000	-4.9
Aroclor-1260-5	6.225	6.126	6.326	464.010	500.000	-7.2
Decachlorobiphenyl	7.535	7.436	7.636	48.140	50.000	-3.7
Tetrachloro-m-xylene	2.778	2.678	2.878	47.380	50.000	-5.2