

Surrogate Summary

SDG No.: Q2893

Client: Alliance Technical Group, LLC - Newark

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PQ070814.D	PIBLK-PQ070814.D	Tetrachloro-m-xyl	1	20	17.3	86		60	140
		Decachlorobiphen	1	20	18.8	94		60	140
		Tetrachloro-m-xyl	2	20	16.6	83		60	140
		Decachlorobiphen	2	20	18.3	91		60	140
PB169530BL	PB169530BL	Tetrachloro-m-xyl	1	20	20.9	104		30	173
		Decachlorobiphen	1	20	22.7	113		10	173
		Tetrachloro-m-xyl	2	20	20.0	100		30	173
		Decachlorobiphen	2	20	22.1	111		10	173
Q2893-07	LOD-MDL-WATER-01-QT3-202	Tetrachloro-m-xyl	1	20	16.3	81		30	173
		Decachlorobiphen	1	20	18.0	90		10	173
		Tetrachloro-m-xyl	2	20	15.6	78		30	173
		Decachlorobiphen	2	20	17.5	87		10	173
I.BLK-PQ070825.D	PIBLK-PQ070825.D	Tetrachloro-m-xyl	1	20	17.3	86		60	140
		Decachlorobiphen	1	20	19.0	95		60	140
		Tetrachloro-m-xyl	2	20	16.3	82		60	140
		Decachlorobiphen	2	20	18.4	92		60	140
I.BLK-PQ070887.D	PIBLK-PQ070887.D	Tetrachloro-m-xyl	1	20	17.6	88		60	140
		Decachlorobiphen	1	20	19.4	97		60	140
		Tetrachloro-m-xyl	2	20	16.4	82		60	140
		Decachlorobiphen	2	20	18.5	92		60	140