

Surrogate Summary

SDG No.: **Q2893**

Client: **Alliance Technical Group, LLC - Newark**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD089961.D	PIBLK-PD089961.D	Decachlorobiphen	1	20	19.3	96		57	171
		Tetrachloro-m-xyl	1	20	17.4	87		61	148
		Decachlorobiphen	2	20	18.3	91		57	171
		Tetrachloro-m-xyl	2	20	17.5	88		61	148
I.BLK-PD090140.D	PIBLK-PD090140.D	Decachlorobiphen	1	20	20.4	102		57	171
		Tetrachloro-m-xyl	1	20	20.2	101		61	148
		Decachlorobiphen	2	20	21.1	105		57	171
		Tetrachloro-m-xyl	2	20	22.1	111		61	148
Q2893-01	LOD-MDL-SOIL-01-QT3-2025	Decachlorobiphen	1	20	18.2	91		20	144
		Tetrachloro-m-xyl	1	20	17.0	85		19	148
		Decachlorobiphen	2	20	19.0	95		20	144
		Tetrachloro-m-xyl	2	20	19.2	96		19	148
		Decachlorobiphen	1	20	18.4	92		20	144
		Tetrachloro-m-xyl	1	20	17.0	85		19	148
		Decachlorobiphen	2	20	19.4	97		20	144
		Tetrachloro-m-xyl	2	20	19.2	96		19	148
I.BLK-PD090152.D	PIBLK-PD090152.D	Decachlorobiphen	1	20	21.4	107		57	171
		Tetrachloro-m-xyl	1	20	20.0	100		61	148
		Decachlorobiphen	2	20	22.9	114		57	171
		Tetrachloro-m-xyl	2	20	22.1	110		61	148
I.BLK-PD090237.D	PIBLK-PD090237.D	Decachlorobiphen	1	20	19.4	97		57	171
		Tetrachloro-m-xyl	1	20	16.6	83		61	148
		Decachlorobiphen	2	20	21.1	106		57	171
		Tetrachloro-m-xyl	2	20	20.3	101		61	148
Q2893-01	LOD-MDL-SOIL-01-QT3-2025	Decachlorobiphen	1	20	20.2	101		20	144
		Tetrachloro-m-xyl	1	20	16.9	85		19	148
		Decachlorobiphen	2	20	22.2	111		20	144
		Tetrachloro-m-xyl	2	20	20.6	103		19	148
		Decachlorobiphen	1	20	20.1	101		20	144
		Tetrachloro-m-xyl	1	20	16.3	81		19	148
		Decachlorobiphen	2	20	21.6	108		20	144
		Tetrachloro-m-xyl	2	20	20.4	102		19	148
I.BLK-PD090252.D	PIBLK-PD090252.D	Decachlorobiphen	1	20	19.7	98		57	171
		Tetrachloro-m-xyl	1	20	16.1	81		61	148
		Decachlorobiphen	2	20	21.8	109		57	171
		Tetrachloro-m-xyl	2	20	20.1	101		61	148