

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

ProjectID: LOD-LOQ-MDL Study

Lab Code: ACE

SDG No.: Q2893

Calibration Sequence : FF080625		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	187893994	110526	FF016188.D
850	88429532	104035	FF016189.D
340	37118669	109173	FF016190.D
170	20282466	119309	FF016191.D
85	10983314	129215	FF016192.D
AVG RF : 114452		% RSD : 8.664	AVG RT : 15.0834