

**TPH GC CONTINUING CALIBRATION SUMMARY**

**50 PPM TRPH STD**

Lab Name: Alliane Contract: ALLI03  
ProjectID: LOD-LOQ-MDL Study  
Lab Code: ACE SDG No.: Q2893  
DataFile: FF016219.D Analyst Name: YP\AJ Analyst Date: 08-26-2025

| Conc. (PPM) | Area Count | RF     | Average RF | %D   |
|-------------|------------|--------|------------|------|
| 850         | 99726234   | 117325 | 114452     | 2.51 |