



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q2893

Continuing Calib Date: 10/07/2025

Initial Calibration Date(s): 10/07/202510/07/2025

Continuing Calib Time: 12:29

Initial Calibration Time(s): 08:5711:09

GC Column: ZB-MR1ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DBCP	5.95	5.96	5.86	6.06	0.01
EDB	2.60	2.61	2.51	2.71	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ALLI03

Lab Code: ACE

SDG NO.: Q2893

Continuing Calib Date: 10/07/2025

Initial Calibration Date(s): 10/07/2025 10/07/2025

Continuing Calib Time: 12:29

Initial Calibration Time(s): 08:57 11:09

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DBCP	4.98	4.98	4.88	5.08	0.00
EDB	1.97	1.97	1.87	2.07	0.00