

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/07/2025 10/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/07/2025
Lab Sample No.: M8011.504.1 0.1 PF **Data File :** PQ070961.D **Time Analyzed:** 13:57

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
DBCP	5.951	5.855	6.055	0.100	0.100	0.0
EDB	2.603	2.507	2.707	0.110	0.100	0.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ALLI03
Lab Code: ACE **SDG NO.:** Q2893
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/07/2025 10/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/07/2025
Lab Sample No.: M8011.504.1 0.1 PF **Data File :** PQ070961.D **Time Analyzed:** 13:57

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
DBCP	4.984	4.881	5.081	0.100	0.100	0.0
EDB	1.974	1.872	2.072	0.110	0.100	0.0