

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ100725

Review By	yogesh	Review On	10/7/2025 2:41:30 PM
Supervise By	mohammad	Supervise On	10/9/2025 1:41:11 AM
SubDirectory	PQ100725	HP Acquire Method	HP Processing Method PQ100725
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24970,PP24971,PP24972,PP24973,PP24974		
CCC Internal Standard/PEM	PP24976,PP24977		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24975		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PQ070946.D	07 Oct 2025 08:17	YP\AJ	Ok
2	M8011.504.1 0.5 PPB ICC	PQ070947.D	07 Oct 2025 08:57	YP\AJ	Ok,M
3	M8011.504.1 0.25 PPB ICC	PQ070948.D	07 Oct 2025 09:07	YP\AJ	Ok
4	M8011.504.1 0.1 PPB ICC	PQ070949.D	07 Oct 2025 10:30	YP\AJ	Ok
5	M8011.504.1 0.05 PPB ICC	PQ070950.D	07 Oct 2025 10:43	YP\AJ	Ok
6	M8011.504.1 0.025 PPB ICC	PQ070951.D	07 Oct 2025 11:09	YP\AJ	Ok,M
7	M8011.504.1 0.1 PPB ICV	PQ070952.D	07 Oct 2025 11:29	YP\AJ	Ok,M
8	RT CHECK	PQ070953.D	07 Oct 2025 12:19	YP\AJ	Ok
9	M8011.504.1 0.1 PPB CCC	PQ070954.D	07 Oct 2025 12:29	YP\AJ	Ok
10	Low-Level LFB-MDL Check	PQ070955.D	07 Oct 2025 12:38	YP\AJ	Ok,M
11	PB170012BL	PQ070956.D	07 Oct 2025 12:48	YP\AJ	Ok
12	PB170012BS	PQ070957.D	07 Oct 2025 12:58	YP\AJ	Ok
13	PB170012BSD	PQ070958.D	07 Oct 2025 13:09	YP\AJ	Ok
14	Q3015-20	PQ070959.D	07 Oct 2025 13:18	YP\AJ	Dilution
15	Q3015-20DL	PQ070960.D	07 Oct 2025 13:47	YP\AJ	Ok
16	M8011.504.1 0.1 PPB CCC	PQ070961.D	07 Oct 2025 13:57	YP\AJ	Ok
17	Q2893-07	PQ070962.D	07 Oct 2025 14:07	YP\AJ	Ok,M
18	Q2893-08	PQ070963.D	07 Oct 2025 14:21	YP\AJ	Ok,M
19	M8011.504.1 0.1 PPB CCC	PQ070964.D	07 Oct 2025 14:31	YP\AJ	Ok

M : Manual Integration