

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2894
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113074.D **Time Analyzed:** 09:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.745	4.648	4.848	557.950	500.000	11.6
Aroclor-1016-2	4.764	4.666	4.866	569.590	500.000	13.9
Aroclor-1016-3	4.821	4.724	4.924	563.830	500.000	12.8
Aroclor-1016-4	4.939	4.843	5.043	536.000	500.000	7.2
Aroclor-1016-5	5.197	5.099	5.299	562.330	500.000	12.5
Aroclor-1260-1	6.233	6.136	6.336	573.260	500.000	14.7
Aroclor-1260-2	6.423	6.326	6.526	542.840	500.000	8.6
Aroclor-1260-3	6.788	6.691	6.891	473.310	500.000	-5.3
Aroclor-1260-4	7.047	6.951	7.151	440.650	500.000	-11.9
Aroclor-1260-5	7.291	7.194	7.394	520.410	500.000	4.1
Decachlorobiphenyl	8.680	8.582	8.782	44.730	50.000	-10.5
Tetrachloro-m-xylene	3.662	3.563	3.763	55.630	50.000	11.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2894
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113074.D **Time Analyzed:** 09:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.728	4.629	4.829	530.680	500.000	6.1
Aroclor-1016-2	4.746	4.647	4.847	520.840	500.000	4.2
Aroclor-1016-3	4.920	4.822	5.022	508.920	500.000	1.8
Aroclor-1016-4	4.963	4.864	5.064	492.130	500.000	-1.6
Aroclor-1016-5	5.175	5.076	5.276	532.190	500.000	6.4
Aroclor-1260-1	6.202	6.104	6.304	506.320	500.000	1.3
Aroclor-1260-2	6.390	6.292	6.492	504.370	500.000	0.9
Aroclor-1260-3	6.542	6.443	6.643	480.850	500.000	-3.8
Aroclor-1260-4	7.011	6.912	7.112	476.380	500.000	-4.7
Aroclor-1260-5	7.253	7.155	7.355	458.050	500.000	-8.4
Decachlorobiphenyl	8.626	8.528	8.728	48.290	50.000	-3.4
Tetrachloro-m-xylene	3.655	3.555	3.755	52.450	50.000	4.9