

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2894
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113089.D **Time Analyzed:** 15:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.745	4.648	4.848	549.720	500.000	9.9
Aroclor-1016-2	4.764	4.666	4.866	557.770	500.000	11.6
Aroclor-1016-3	4.821	4.724	4.924	560.020	500.000	12.0
Aroclor-1016-4	4.940	4.843	5.043	521.460	500.000	4.3
Aroclor-1016-5	5.197	5.099	5.299	588.170	500.000	17.6
Aroclor-1260-1	6.233	6.136	6.336	578.100	500.000	15.6
Aroclor-1260-2	6.423	6.326	6.526	553.590	500.000	10.7
Aroclor-1260-3	6.789	6.691	6.891	548.020	500.000	9.6
Aroclor-1260-4	7.048	6.951	7.151	553.140	500.000	10.6
Aroclor-1260-5	7.291	7.194	7.394	559.300	500.000	11.9
Decachlorobiphenyl	8.680	8.582	8.782	53.560	50.000	7.1
Tetrachloro-m-xylene	3.662	3.563	3.763	55.020	50.000	10.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2894
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113089.D **Time Analyzed:** 15:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.728	4.629	4.829	525.810	500.000	5.2
Aroclor-1016-2	4.747	4.647	4.847	528.740	500.000	5.7
Aroclor-1016-3	4.921	4.822	5.022	517.620	500.000	3.5
Aroclor-1016-4	4.964	4.864	5.064	497.210	500.000	-0.6
Aroclor-1016-5	5.176	5.076	5.276	552.910	500.000	10.6
Aroclor-1260-1	6.203	6.104	6.304	512.480	500.000	2.5
Aroclor-1260-2	6.391	6.292	6.492	509.070	500.000	1.8
Aroclor-1260-3	6.543	6.443	6.643	478.950	500.000	-4.2
Aroclor-1260-4	7.012	6.912	7.112	498.110	500.000	-0.4
Aroclor-1260-5	7.254	7.155	7.355	490.130	500.000	-2.0
Decachlorobiphenyl	8.627	8.528	8.728	51.510	50.000	3.0
Tetrachloro-m-xylene	3.655	3.555	3.755	53.390	50.000	6.8