

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2894
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113104.D **Time Analyzed:** 21:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.747	4.648	4.848	546.270	500.000	9.3
Aroclor-1016-2	4.765	4.666	4.866	558.770	500.000	11.8
Aroclor-1016-3	4.822	4.724	4.924	561.700	500.000	12.3
Aroclor-1016-4	4.941	4.843	5.043	508.020	500.000	1.6
Aroclor-1016-5	5.198	5.099	5.299	599.770	500.000	20.0
Aroclor-1260-1	6.234	6.136	6.336	545.420	500.000	9.1
Aroclor-1260-2	6.424	6.326	6.526	523.590	500.000	4.7
Aroclor-1260-3	6.790	6.691	6.891	503.040	500.000	0.6
Aroclor-1260-4	7.049	6.951	7.151	496.540	500.000	-0.7
Aroclor-1260-5	7.292	7.194	7.394	500.680	500.000	0.1
Decachlorobiphenyl	8.680	8.582	8.782	47.900	50.000	-4.2
Tetrachloro-m-xylene	3.662	3.563	3.763	55.180	50.000	10.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2894
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/20/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113104.D **Time Analyzed:** 21:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.728	4.629	4.829	525.020	500.000	5.0
Aroclor-1016-2	4.746	4.647	4.847	522.860	500.000	4.6
Aroclor-1016-3	4.921	4.822	5.022	513.260	500.000	2.7
Aroclor-1016-4	4.964	4.864	5.064	496.610	500.000	-0.7
Aroclor-1016-5	5.175	5.076	5.276	557.720	500.000	11.5
Aroclor-1260-1	6.203	6.104	6.304	490.130	500.000	-2.0
Aroclor-1260-2	6.391	6.292	6.492	488.970	500.000	-2.2
Aroclor-1260-3	6.542	6.443	6.643	453.790	500.000	-9.2
Aroclor-1260-4	7.012	6.912	7.112	473.090	500.000	-5.4
Aroclor-1260-5	7.254	7.155	7.355	471.190	500.000	-5.8
Decachlorobiphenyl	8.626	8.528	8.728	50.940	50.000	1.9
Tetrachloro-m-xylene	3.654	3.555	3.755	53.170	50.000	6.3